



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1363

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Krohne Messtechnik GmbH
Address	Ludwig-Krohne-Stral 5 47058 Duisburg, Germany
Project Number	LR1363-2
Product	Variable area Flowmeter
Model Number	DK32/R1/S, DK34/R1/L, DK34/K1/L, DK32/ESK,
Ratings	DK32/R1/S, 30Vdc, 0.5A DK34/R1/L, 30Vdc, 0.5A DK34/K1/L, 8Vdc, 1mA DK32/ESK, 32Vdc, 4-20mA
Applicable Standards	<ul style="list-style-type: none">- CSA C22.2 No. 61010-1 - Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016- UL61010-1 - Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Including Revisions through July 19, 2019
Factory/Manufacturing Location	Same as above

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Andrew Saunders

Signature:

Date: May 22, 2020