



EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert NO.GYJ12.1131X

This is to certify that the product

All-metal Miniature Flowmeter

manufactured by **KROHNE Messtechnik GmbH**

(Address:Ludwig - Krohne Strasse 5, Duisburg, Germany)

which model is **DK32 Series, DK34 Series**

Ex marking **Ex nA II C T1~T6 Gc**

product standard /

drawing number **ZZ 8144800100b**

has been inspected and certified by NEPSI, and that it conforms
to GB 3836.1-2010,GB 3836.8-2003

This Approval shall remain in force until **2017.07.01**

Remarks

- 1.Conditions for safe use are specified in the attachment to this certificate.
- 2.Symbol "X" placed after the certification number denotes specific conditions of use, which are specified in the attachment to this certificate.
- 3.Model designation is specified in the attachment to this certificate.

Director

National Supervision and Inspection Centre for
Explosion Protection and Safety of Instrumentation

Issued Date **2012.07.02**

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel:0086 21 64368180
Fax:0086 21 64844580

国家级仪器仪表防爆安全监督检验站

National Supervision and Inspection Centre for Explosion Protection and Safety of Instrumentation

(GYJ12.1131X)

(Attachment I)

Attachment I GYJ12.1131X

DK32 Series & DK34 Series All-metal Miniature Flowmeter, manufactured by Krohne Meßtechnik GmbH & Co KG, has been certified National Supervision and Inspection Center for Explosion Protection and Safety of Instrumentation (NEPSI).

All-metal Miniature Flowmeter accords with following standards:

GB 3836.1-2010 "Explosive atmospheres Part 1: Equipment-General requirements"

GB 3836.8-2003 "Electrical apparatus for explosive gas atmospheres Part 8: Type of protection 'n'"

All-metal Miniature Flowmeter has the Ex-marking Ex nA II T1~T6 Gc.

Following products are covered by this certificate.

DK \bar{a} / \bar{b} / \bar{c} / L / A

\bar{a} : 32, 34

\bar{b} : RE, RA

\bar{c} : K1min, K1max, K2

1. Specific condition for safe use

The suffix "X" denotes Only proximity switch typed SC2-NO-Y or SJ2-SN or SJ2-S1N from Germany PEPPERL+FUCHS could be installed in this all-metal Miniature Flowmeter. And the rated voltage of the proximity switch is 20V. External provision should be made to protect the power supply exceeding 40% of the rated voltage of the apparatus.

2. Condition for safe use

2.1 The relation among temperature class, ambient temperature and maximum temperature of process medium is as following.

Temperature code	Ambient temperature	Maximum temperature of process medium	
		DK32 Series	DK34 Series
T6	(-20~+40) °C	75°C	80°C
T6	(-20~+50) °C	70°C	70°C
T5	(-20~+40) °C	100°C	100°C
T5	(-20~+50) °C	95°C	100°C
T5	(-20~+60) °C	85°C	90°C
T4	(-20~+40) °C	135°C	135°C
T4	(-20~+50) °C	130°C	135°C
T4	(-20~+60) °C	120°C	130°C
T1~T3	(-20~+40) °C	135°C	150°C
T1~T3	(-20~+50) °C	130°C	140°C
T1~T3	(-20~+60) °C	120°C	130°C

2.2 End users is not permitted to change any components insides.

2.3 When installation, use and maintenance of All-metal Miniature Flowmeter, observe following standards.

GB3836.13-1997“Electrical apparatus for explosive gas atmospheres Part 13:Repair and overhaul for apparatus used in explosive gas atmospheres”

GB3836.15-2000“Electrical apparatus for explosive gas atmospheres Part 15:Electrical installations in hazardous area (other than mines)”

GB 3836.16-2006 “Electrical apparatus for explosive gas atmospheres – Part 16: Inspection and maintenance of electrical installation in hazardous areas (other than mines)

GB 50257:1996 ”Code for construction and acceptance of electric device for explosion atmospheres and fire hazard electrical equipment installation engineering”

3. Manufacturer’s Responsibility

3.1 Special condition for safe use specified above should be included in the instruction manual.

3.2 Manufacturing should be done according to the documentation approved by NEPSI.

3.3 Any modification with influence on the type of protection should be submitted to NEPSI before application.

3.4 Following items should be added to the nameplate

- a) NEPSI log 
- b) Ex marking
- c) Number of certificate
- d) Ambient temperature range

**National Supervision and Inspection Center
for Explosion Protection and Safety of Instrumentation**

July 2, 2012