

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

KROHNE Oil & Gas  
Minervum 7441  
4817 ZG Breda  
The Netherlands

Measuring instrument

An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : Summit 8800  
Manufacturer's mark or name : KROHNE Oil & Gas  
Conversion principle : PTZ  
Device type : 2 (separate component)  
Ambient temperature range : -25 °C / +55 °C (HART input signals)  
-10 °C / +55 °C (analogue input signals)  
Designed for : non-condensing humidity  
Environment classes : M2 and M3 / E2 and E3  
+ The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T10185 revision 19;
- Documentation folder T10185-4.

Valid until

3 January 2029

Remarks

This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
25 March 2019

  
C. Oosterman  
Head Certification Board